

Product Quality Control Report
By Nantong Egens Biotechnology Limited Company

Commodity Name	One Step AMP/BAR/COC/MDMA/THC/BZO Rapid Screen Test (Colloidal Gold)		
Lot Number	20150402	Model	cassette
Quantity	1000	Manufacture Date	20150402
Examining Date	20150402	Expiry Date	20170401
Examining Criteria	Regulations of Company's Quality Control		
Examining Items	Technical Requirements	Test Result	Determinant
Appearance	The appearance should be even, the materials are hard adhered and the content is complete.	Good	Pass
Physical Inspection	The move speed of liquid is not less than 10 mm/min.	Good	Pass
Sensitivity	To use AMP 1000ng/ml BAR 300 ng/ml COC 300ng/ml MDMA 500ng/ml THC 50ng/ml BZO 150ng/ml standard product for inspection, within five minutes.	+	Pass
Reproducibility (10 reference samples)	Identical reactivity	Good	Pass
Conclusion:			
This batch of products are qualified.			
Tested by:	Zhou Jie	Date:2015.04.02	
Reviewed by:	Lu Yanzi	Date:2015.04.02	
Approved by:	Zhang Juan	Date:2015.04.02	